



ALFA GEOTECH CO., LTD.

Calibration Sheet : Pressure Transducer

Summary Calibration Load Cell

New

Old

Date	Pressure No.	Calibration Factor		r ²	Date	Calibration Factor		r ²	Difference Ratio (%)	Remark
		ksc / _(V/V)				ksc / _(V/V)				
Aug-16	1	-		-					-	Broken
Aug-16	2	-		-					-	Broken
Aug-16	3									
Aug-16	4	-		-		2,068.7130			-	Broken
Aug-16	5	2,065.5534	202,564.8136	0.99993		2,073.1985			0.37	
Aug-16	6	2,073.2647	203,321.0412	0.99995		2,052.2716			1.02	
Aug-16	7	2,074.6669	203,458.5531	0.99994		2,062.0244			0.61	
Aug-16	8	2,066.2953	202,637.5690	0.99994		2,054.9569			0.55	
Aug-16	9	2,049.0425	200,945.6237	0.99993		2,040.3367			0.43	
Aug-16	10	2,041.9727	200,252.3021	0.99996		2,025.1598			0.83	
Aug-16	11	2,038.2596	199,888.1633	0.99997		2,020.4440			0.88	
Aug-16	12	2,055.9956	201,627.4990	0.99996		2,034.5562			1.05	
Aug-16	13	2,031.0359	199,179.7469	0.99997		2,040.0565			0.44	
Aug-16	14	2,075.6059	203,550.6387	0.99993						
Aug-16	15	-		-					-	Broken
Aug-16	16	2,060.8605	202,104.5852	0.99994						
Aug-16	18	2,045.7722	200,624.9076	0.99993						
Aug-16	21					2,055.0419				
Aug-16	22	2,050.2962	201,068.5649	0.99994		2,051.1895			0.04	
Aug-16	29	2,071.7211	203,169.6711	0.99993		2,060.6014			0.54	
Aug-16	30	2,079.8280	203,964.6941	0.99993		2,070.1848			0.47	
Aug-16	31	2,062.9861	202,313.0470	0.99993		2,049.5537			0.66	
Aug-16	33	2,046.6303	200,709.0663	0.99995		2,044.7977			0.09	
Aug-16	36	2,042.8633	200,339.6420	0.99994		2,040.5948			0.11	
Aug-16	37	-		-					-	Broken
Aug-16	38					2,066.5863				